



## **Declaration of Conformity**

For the following equipment :

Product Name: Switching Power Supply

Model Designation: X-05-Y (X=NFM,PM) (Y=3.3,5,12,15,24)

is herewith confirmed to comply with the requirements set out in the Council Directive, the following standards were applied :

## RoHS Directive (2011/65/EU)

MDD Directive (93/42/EEC)

EN60601-1:2006+A11+A12

TUV certificate No : TA50220068

## **EMI (Electro-Magnetic Interference)**

EMI (Electro-Magnetic Interference)						
Conducted emissior	n / Radiated emission					
	EN55011:2009+A1:2010 (Group	,				
	EN55022:2010	Class B				
	EN61000-6-3:2007+A1:2011+AC	C:2012				
Harmonic current	EN61000-3-2:2014					
Voltage flicker	EN61000-3-3:2013					
EMS (Electro-Magnetic Susceptibility)						
EN60601-1-2:2015	EN55024:2010+A1:2015 EN61204-3:2	2000 EN61000-6-2:2005+AC:2005				
ESD air	EN61000-4-2:2009	Level 4 15KV				
ESD contact	EN61000-4-2:2009	Level 4 8KV				

EN61000-4-3:2006+A1:2008+A2:2010	Level 3	10V/m(80MHz-2.7GHz)
EN61000-4-3:2006+A1:2008+A2:2010	Table 9	RF wireless communications equipment
EN61000-4-4:2012	Level 3	2KV/100KHz
EN61000-4-5:2014	Level 4	2KV/Line-Line
EN61000-4-6:2014	Level 3	10V
EN61000-4-8:2010	Level 4	30A/m
	EN61000-4-3:2006+A1:2008+A2:2010 EN61000-4-4:2012 EN61000-4-5:2014 EN61000-4-6:2014	EN61000-4-3:2006+A1:2008+A2:2010 Table 9   EN61000-4-4:2012 Level 3   EN61000-4-5:2014 Level 4   EN61000-4-6:2014 Level 3

## Note:

Voltage dip, interruption

A component power supply will be installed into final equipmen. Since EMC performance will be affected by the complete installation, the final equipment manufacturers must re-qualify EMC Directive on the complete installation again. For guidance on how to perform these EMC tests, please refer to TDF (Technical Documentation File)..

EN61000-4-11:2004 100% dip 1 periods 30% dip 25 periods 100% interruptions 250 periods

This Declaration is effective from serial number EB6xxxxxxx

Person responsible for marking this declaration :

MEAN WELL Enterprises Co., Ltd.							
(Manufacturer Name) No.28, Wuguan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan							
No.20, Wuquan Siu Ku., Wugu Disi., New Taiper City 240, Taiwan							
(Manufacturer Address)	1		MIO				
Johnny Huang/Manager, Certification Center	: Tolen	Ted Cheng/Director, Sales Dept. :	Jed Chang				
(Name / Position)	(Signature)	(Name / Position)	(Signature)				
			2				
Taiwan	Sep. 30th, 2016						
(Place)	(Date)						